



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

Pressure Test Certificate

Test Date: 11/10/14

Description: UCTD

Sensor Information:

Model Number: UCTD

Serial Number: 0211

Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For: 15 Minutes

High Pressure Test: 2900 PSI Held For: 15 Minutes

Passed Test: Yes

Tested By: EM

High pressure is
generally equal
to the maximum
depth rating of
the instrument

